

INDUCTANCE DECADE BOX

Model : LBOX-405

ISO-9001, CE, IEC1010

FEATURES

* Applications :

General applications
Troubleshooting, maintenance
Education and Vocational training
Production line testing
Radio and TV services
Working standards
Research design & develop
Physics laboratory work

* Pocket size, offering accurate, reliable performance.


* 10 uH to 111.1 mH, wide range and high resolution (10 uH per step), practical and versatile tools.

* With four decades of inductance.

* Slide switches that allow the user to simply add or subtract for desired value.

* Terminals with multi way binding posts, one to switch shield case.

* ABS plastic housing case, rugged components.



勝特力材料 886-3-5753170
勝特力电子(上海) 86-21-34970699
勝特力电子(深圳) 86-755-83298787
[Http://www.100y.com.tw](http://www.100y.com.tw)

Lutron

LUTRON ELECTRONIC

The Art of Measurement

INDUCTANCE DECADE BOX

Model : LBOX-405

FEATURES	
* Applications : <i>General applications</i> <i>Troubleshooting, maintenance</i> <i>Education and Vocational training</i> <i>Production line testing</i> <i>Radio and TV services</i> <i>Working standards</i> <i>Research design & develop</i> <i>Physics laboratory work</i>	* 10 uH to 111.1 mH, wide range and high resolution (10 uH per step), practical and versatile tools. * With four decades of inductance. * Slide switches that allow the user to simply add or subtract for desired value. * Terminals with multi way binding posts, one to switch shield case. * ABS plastic housing case, rugged components.
* Pocket size, offering accurate, reliable performance.	

SPECIFICATIONS	
Range	10 uH to 111.1 mH (10 uH per step)
Accuracy	5% inductors used throughout. @ 23 ± 5°C @ 1 KHz test frequency
Max. Rating Current	10 mA DC or AC.
Internal Residual Inductance	Approx. 0.5 uH.
Power Supply	None.
Operating Temperature	0 °C to 50 °C (32 °F to 122 °F).
Operating Humidity	Less than 80% RH.
Weight	333 g/0.74 lb.
Dimension	14.7 cm x 11.7 cm x 5.5 cm. (5.79 x 4.61 x 2.16 inch).
Accessories	Operation Manual..... 1 PC.

勝特力材料 886-3-5753170
 勝特力电子(上海) 86-21-34970699
 勝特力电子(深圳) 86-755-83298787
[Http://www.100y.com.tw](http://www.100y.com.tw)

TESTING PROCEDURE
1) Start with all switches up (OUT) for min. inductance. 2) Switch down (IN) to add Inductance value. 3) The " Ground Terminal " is connected to the metal enclosure of all switches For some special application may connect the " Ground Terminal " to the external equipment to prevent other environment interference.

* Appearance and specifications listed in this brochure are subject to change without notice. 0404-LBOX405