

g

GE Lighting

***GE Specialty
medical, dental
and scientific lamps
can be instrumental
in your work***

勝特力材料 886-3-5753170
勝特力电子(上海) 86-21-54151736
勝特力电子(深圳) 86-755-83298787
[Http://www.100y.com.tw](http://www.100y.com.tw)



Examine GE's full line of medical, dental and scientific lamps.

From the microscope to the operating room, from dental equipment to laboratory instruments, from fiber optics to measuring and inspection equipment, GE has the right lamps

for your applications. All of the broad range of GE lamps have one thing in common. Each is engineered to GE's high standards of quality and precision for excellent value.



勝特力材料 886-3-5753170
 勝特力电子(上海) 86-21-54151736
 勝特力电子(深圳) 86-755-83298787
[Http://www.100y.com.tw](http://www.100y.com.tw)

| Description | Volts | GE Prod. Code | Bulb | Base | Watts | Std. Pkg. Qty. | Additional Information | Fila-ment Design | MOL (in.) | Avg. LCL (in.) | Rated Life hours | App. Initial Lumens |
|----------------------|---------|---------------|-------|----------------|-------|----------------|-------------------------|------------------|-----------|----------------|------------------|---------------------|
| PROJECTION | | | | | | | | | | | | |
| BLC | 115-125 | 29140 | S-11 | D.C. Bay. | 30 | 120 | Biascope/Microscope | 2775 | CC-2V | 2.375 | 1.375 | 400 |
| CAX | 115/5 | 29171 | T-8 | D.C. Bay. | 50 | 24 | Optical Projection | 2875 | CC-13 | 3.125 | 1.375 | 775 |
| CAX | 130 | 29169 | T-8 | D.C. Bay. | 50 | 24 | Optical Projection | 2875 | CC-13 | 3.125 | 1.375 | 775 |
| DDL | 20 | 43537 | MR-16 | 2-Pin | 150 | 20 | Microfilm | 3150 | CC-6 | 1.75 | | 500 |
| DNE | 120 | 40161 | MR-16 | 2-Pin Vented | 750 | 24 | Fiber Optics Projection | 3400 | CC-6 | 1.774 | 0.625 | 25 |
| DNF | 21 | 39742 | MR-16 | 2-Pin Vented | 150 | 24 | Fiber Optics Projection | 3400 | CC-6 | 1.744 | 0.625 | 25 |
| DZA | 10.8 | 37346 | T-3.5 | Min 2-Pin | 30 | 24 | Fiber Optics Projection | 3100 | C-6 | 2 | 1.062 | 400 |
| DZE/FDS | 24 | 37695 | T-4 | 2-Pin Prefocus | 150 | 24 | Fiber Optics Projection | 3250 | C-6 | 2.867 | 1.312 | 100 |
| EFM | 8 | 41251 | MR-16 | Glass 2-Pin | 50 | 24 | Fiber Optics Projection | 3400 | CC-6 | 1.774 | 0.625 | 25 |
| EFF | 12 | 41253 | MR-16 | Glass 2-Pin | 100 | 24 | Endoscope | 3350 | CC-6 | 1.75 | 1.656 | 50 |
| EFR | 15 | 41254 | MR-16 | Glass 2-Pin | 150 | 24 | Biascope/Microscope | 3350 | CC-6 | 1.75 | 1.656 | 50 |
| EHJ | 24 | 14874 | T-4 | Glass 2-Pin | 250 | 100 | Medical Testing | 3400 | C-6 Oval | 2.25 | 1.375 | 50 |
| EJA | 21 | 32882 | MR-16 | 2-Pin | 150 | 24 | Fiber Optics Projection | 3350 | CC-6 | 1.75 | - | 25 |
| EJM | 21 | 29151 | MR-16 | 2-Pin | 150 | 24 | Fiber Optics Projection | 3350 | CC-6 | 1.75 | - | 40 |
| EJV | 21 | 32831 | MR-16 | 2-Pin | 150 | 24 | Fiber Optics Projection | 3350 | CC-6 | 1.75 | - | 40 |
| EJY | 19 | 32886 | MR-16 | 2-Pin | 80 | 24 | Fiber Optics Projection | 3350 | CC-6 | 1.75 | - | 25 |
| EKE | 21.5 | 35200 | MR-16 | 2-Pin | 150 | 24 | Fiber Optics Projection | 3250 | CC-6 | 1.75 | - | 200 |
| EKP/ENA | 30 | 35800 | MR-16 | 2-Pin | 80 | 24 | Fiber Optics Projection | 3350 | CC-6 | 1.75 | - | 25 |
| EKZ | 10.8 | 36902 | MR-16 | 2-Pin | 30 | 24 | Fiber Optics Projection | 3400 | C-6 | 1.75 | - | 200 |
| EML | 24 | 42612 | T-4 | Min. 2-Pin | 175 | 24 | Fiber Optics Projection | 3200 | C-6 | 2.125 | 1.062 | 125 |
| ENW/ENC | 19 | 40248 | MR-16 | 2-Pin | 80 | 24 | Fiber Optics Projection | 3200 | CC-6 | 1.75 | - | 200 |
| ENX | 82 | 41705 | MR-16 | Oval 2-pin | 360 | 20 | Overhead projection | 3300 | CC-6 | 1.75 | | 75 |
| EPN | 12 | 41430 | MR-16 | 2-Pin | 35 | 20 | 8mm projection | 3300 | CC-6 | 1.75 | | 50 |
| EPT | 10.8 | 41729 | MR-16 | 2-Pin | 42 | 24 | Fiber Optics Projection | 2900 | C-6 | 1.75 | - | 8000 |
| ETJ | 120 | 11322 | MR-16 | Oval 2-Pin | 250 | 24 | Fiber Optics Projection | 3250 | CC-8 | 1.75 | - | 175 |
| EXR | 82 | 12092 | MR-13 | 2-pin | 300 | 20 | Slide projection | 3350 | CC-8 | | | 35 |
| EYB | 82 | 12696 | T3.5 | Min-2-pin | 360 | 24 | | 3300 | CC-8 | 2.25 | 1.25 | 75 |
| EZM | 37.5 | 29469 | MR-24 | Special 2-Pin | 300 | 4 | Endoscope | 5500 | Arc Lamp | - | - | 25 |
| FCS | 24 | 13598 | T-4 | Glass 2-Pin | 150 | 100 | Medical Testing | 3400 | C-6 Oval | 2 | 1.187 | 50 |
| FDT | 12 | 35321 | T-3 | 2-Pin Prefocus | 100 | 24 | Fiber Optics Projection | 3300 | C-6 | 2.125 | 1.062 | 50 |
| FDV | 24 | 36878 | T-4 | Glass 2-Pin | 150 | 24 | Blood Analyzer Equip. | 3250 | C-6 Oval | 2 | 1.187 | 100 |
| FXL | 82 | 21613 | MR-16 | GY5.3 | 410 | 20 | Overhead projection | 3300 | CC-8 | 1.75 | | 38 |
| QUARTZ | | | | | | | | | | | | |
| Q20T2.5/12V/CL | 12 | 34715 | T-3 | 1-Pin G4 | 20 | 100 | Clear | | C-6 | 1.25 | .75 | 2000 |
| Q20MR16/C/FL40°-BAB | 12 | 20814 | MR-16 | 2-Pin GX5.3 | 20 | 20 | Flood | 2900 | C-6 | 1.875 | | 5000 |
| Q45T2/1/2/CL | 7 | 23850 | T2.5 | RSC | 45 | 12 | Clear-Instrument | - | C-8 | 2.062 | (137) | 1000 |
| Q50MR16/C/NFL25°-EXZ | 12 | 20835 | MR-16 | 2-Pin GX5.3 | 50 | 20 | Narrow Flood | 3050 | C-6 | 1.875 | | 6000 |
| Q50MR16/C/FL40°-EXN | 12 | 20833 | MR16 | 2-Pin GX5.3 | 50 | 20 | Flood | 3050 | C-6 | 1.875 | | 1700 |
| Q100CL/DC/2V | 120 | 44386 | T-4 | D.C. Bay. | 100 | 6 | Clear ANSI:ESR (101) | - | CC-2V | 2.437 | 1.375 | 750 |
| Q100CL/DC | 120 | 15508 | T-4 | D.C. Bay. | 100 | 6 | Clear | 2950 | CC-8 | 2.438 | 1.375 | 2000 |
| Q100CL/MC | 120 | 15507 | T-4 | Mini-Cand. | 100 | 6 | Clear | 2950 | CC-8 | 2.813 | 1.375 | 2000 |
| Q100CL/MC/2V-ESN | 120 | 44385 | T-4 | Mini-Cand. | 100 | 6 | Clear | 2950 | CC-2V | 2.813 | 1.375 | 750 |
| Q100CL/MC/2V | 120 | 44385 | T-4 | Mini-Can | 100 | 6 | Clear ANSI:ESN (101) | - | CC-2V | 2.812 | 1.375 | 750 |
| Q100DC/2V | 120 | 44657 | T-4 | D.C. Bay. | 100 | 6 | Frosted ANSI:ETD (101) | - | CC-2V | 2.437 | 1.375 | 750 |
| Q150CL/DC/2V | 120 | 44384 | T-4 | D.C. Bay. | 150 | 6 | Clear ANSI:ESP (101) | - | CC-2V | 2.437 | 1.375 | 1000 |
| Q150CL/DC-ETC | 120 | 43693 | T-4 | D.C. Bay. | 150 | 6 | Clear | 2950 | CC-8 | 2.5 | 1.375 | 2000 |
| Q150CL/MC/2V | 120 | 44383 | T-4 | Mini-Can | 150 | 6 | Clear ANSI:ESL(101) | - | CC-2V | 2.812 | 1.375 | 1000 |
| Q150CL/MC-ETG | 120 | 43694 | T-4 | Mini-Cand | 150 | 6 | Clear | 2950 | CC-8 | 3.0 | 1.375 | 2000 |
| Q150T4/CL | 25 | 23710 | T-4 | RSC | 150 | 12 | Clear- Dental Spotlight | - | CC-8 | 2.562 | - | 3000 |
| Q235T4/3 | 33 | 11548 | T-4 | 2-Pin Prefocus | 235 | 12 | Frosted-Instrument | - | CC-6 | 2.5 | 1.531 | (103) |
| Q250CL/MC-EHT | 120 | 43699 | T-4 | Mini-Cand | 250 | 6 | Clear | 2950 | CC-8 | 3.156 | 1.625 | 2000 |
| Q300T3/CL-EHM | 120 | 43703 | T2.5 | R7s | 300 | 144 | Clear, Horizontal | 2950 | C-8 | 4.688 | 2.25 | 2000 |
| Q500T3/CL-FCL | 120 | 23731 | T2.5 | R7s | 500 | 12 | Clear, Horizontal | 3000 | C-8 | 4.688 | 2.25 | 2000 |

| Desc | Volts | GE Prod. Code | Bulb | Base | Watts or Amps | Std. Pkg. Qty. | Additional Information | Fila-ment Design | MOL (in.) | LCL (in.) | Rated Avg. Life hours | Mean Spher. Candle App. |
|------------------|-------------|---------------|--------|----------------|----------------|----------------|---|------------------|-----------|-----------|-----------------------|-------------------------|
| Miniature | | | | | | | | | | | | |
| 44 | 6.3 | 25450 | T3.25 | Min. Bay. | | 50 | | C-2R | 1.19 | .78 | 3000 | |
| 51 | 7.5 | 25529 | G3.5 | Min. Bay. | | 50 | Indicator | C-2R | .94 | .5 | 1000 | |
| 68 | 13.5 | 25692 | G6 | D.C. Bay. | | 50 | | C-2R | 1.44 | .81 | 5000 | |
| 73/BP2 | 14 | 23015 | T1.75 | Wedge | .08A | 10 | Indicator | C-2F | 0.8 | 0.4 | 15,000 | 0.3 |
| 85 | 28 | 40969 | T1.75 | Wedge | | 50 | Indicator | C-2F | .8 | .4 | 7000 | |
| 157 | 5.8 | 25927 | G-6 | Med.Screw | 1.10A | 10 | Medical | C-2R | 1.22 | 1 | 50 | 8.1 |
| 168/BP2 | 14 | 12327 | T3.25 | Wedge | | 48 | | C-2F | 1.06 | .56 | 1500 | |
| 259 | 6.3 | 26099 | T3.25 | Wedge | | 50 | Radio | C-2R | 1.06 | .65 | 5000+ | |
| 311 | 28 | 26191 | S11 | S.C. Bay | | 10 | | C-2V | 2.38 | 1.25 | 300 | |
| 313 | 28 | 26212 | T3.25 | Min. Bay. | | 50 | | C-2F | 1.19 | .62 | 500 | |
| 316 | 6 | 26243 | T3.25 | Min.Bay. | .70A | 10 | Instrument | C-2R | 1.19 | 0.62 | 500 | 3.4 |
| 327 | 28 | 28519 | T1.75 | S.C. Mid Fig | | 50 | | C-2F | .63 | .38 | 4000 | |
| 382 | 14 | 28657 | T1.75 | S.C. Mid Fig | | 50 | Indicator | C-2F | .63 | .38 | 40,000 | |
| 385 | 28 | 28660 | T1.75 | S.C. Mid Fig | | 50 | Indicator | C-2F | .81 | .44 | 10,000 | |
| 387 | 28 | 28664 | T1.75 | S.C. Mid Fig | | 50 | Indicator | C-2F | .63 | .38 | 7000 | |
| 400 | 28 | 38918 | T3.25 | Wedge | | 50 | | C-2F | 1.06 | .56 | 1000 | |
| 656 | 28 | 38866 | T3.25 | Wedge | | 50 | Indicator | C-2F | 1.06 | .56 | 2500 | |
| 657 | 28 | 38196 | T3.25 | Wedge | | 50 | Indicator | C-2F | 1.06 | .56 | 15,000 | |
| 755 | | 26591 | T3.25 | Min Bay | | 50 | Indicator | C-2R | 1.19 | .78 | 20,000 | |
| 756 | 14 | 26593 | T3.25 | Min Bay | | 50 | Indicator | C-2F | 1.19 | .62 | 15,000 | |
| 757 | 28 | 26599 | T3.25 | Min Bay | | 50 | Indicator | C-2F | 1.19 | .62 | 15,000 | |
| 778 | 6 | 49718 | T-2.75 | G-4 2-Pin | 3.33A | 10 | Instrument - halogen High Temp. Fixtures (306) | C-6 | 1.05 | 0.77 | 100 | 32 |
| 787 | 6 | 43115 | T-2.25 | G-4 2-Pin | 1.67A | 10 | Instrument - halogen (306) | C-6 | 1.0 | 0.77 | 100 | 16 |
| 788 | 6 | 43117 | T-2.25 | G-4 2-Pin | 3.33A | 10 | Instrument - halogen (306) | C-6 | 1.0 | 0.77 | 100 | 22 |
| 789 | 12 | 43119 | T-2.25 | G-4 2-Pin | 1.17A | 10 | Instrument - halogen (306) | C-6 | 1.05 | 0.77 | 100 | 32 |
| 790 | 14 | 43121 | T-2.75 | G-4 2-Pin | 1.79A | 10 | Instrument - halogen (306) | C-6 | 1.05 | 0.77 | 200 | 42 |
| 791 | 14 | 43123 | T-2.75 | G-4 2-Pin | 2.05A | 10 | Instrument - halogen (306) | C-6 | 1.05 | 0.77 | 200 | 61 |
| 909 | 6 | 44756 | T5 | Wedge | | 50 | Emergency Lighting | C-2R | 1.49 | .81 | 50 | |
| 1096 | 6 | 37169 | S-8 | D.C. Pf. (S) | 4.5A | 10 | Instr - Microscope Illuminator | Horiz. | 2 | 1 | 500 | 30 |
| 1152 | 12.8 | 26945 | S-8 | D.C. Bay. | 1.34A | 10 | Special Service | C-2R | 2 | 1.25 | 500 | 21 |
| 1156/BP2 | 12.8 | 12344 | S8 | S.C. Bay | | 48 | Auto-Stop | C-6 | 2 | 1.25 | 1200 | |
| 1383 | 13 | 27150 | R12 | S.C. Bay | 20W | 10 | Auto-Reflectionalized | C-8 | 2.63 | | 300 | |
| 1385 | 28 | 27154 | R12 | S.C. Bay | 20W | 10 | Special-Reflectionalized | CC-8 | 2.63 | | 300 | |
| 1434 | 3.7 | 27193 | T-5 | S.C. Bay. | 2.75A | 10 | Instrument, Photocell Exciter - Silver Plated Contact | C-6 | 1.75 | 1.12 | 100 | 11 |
| 1460 | 6.5 | 28310 | S-8 | D.C. Pf. (A) | 2.75A | 10 | Medical | C-6 | 2 | 1.25 | 100 | 23 |
| 1468 | 6 | 27305 | S-8 | D.C. Pf. (S) | 4.5A | 10 | Medical Instrument - Burning position BUTH | C-2R | 2 | 1.25 | 500 | 30 |
| 1489 | 6.5 | 27369 | T-5 | S.C. Bay. | 2.75A | 10 | Instrument (Axial alignment tolerance .016) | C-6 | 1.75 | 1.125 | 125 | 24 |
| 1493 | 6.5 | 27382 | S-8 | D.C. Bay. | 2.75A | 10 | Instrument Microscope Illum. (L.C.L. & A.A. Tol. .047) | C-6 | 2 | 1.12 | 100 | 23 |
| 1561 | 6.3 | 27431 | S-11 | S.C. Pf. (B) | 4.0A | 10 | Instrument, Colorimeter | C-8Z | 2.38 | 0.88 | 1500 | 24 |
| 1594 | 6 | 27439 | S-8 | D.C. Bay. | 5.0A | 10 | Medical Instrument Burning Position BUTH | C-6 | 2 | 1.25 | 250 | 36 |
| 1612 | 5.4 | 27461 | S-8 | D.C. Bay. | 1.90A | 10 | Instrument, Microscope Illum (L.C.L. & A.A. Tol. .047) | C-6 | 2 | 1.25 | 1000 | 10 |
| 1619 | 6.7 | 27472 | S-8 | S.C. Bay. | 1.90A | 10 | Instrument | C-6 | 2 | 1.12 | 500 | 15 |
| 1630 | 6.5 | 27488 | S-8 | D.C. Pf. (A) | 2.75A | 10 | Instrument, Microscope Illum. | C-6 | 2 | 1 | 100 | 23 |
| 1634 | 20 | 27496 | S-8 | D.C. Pf. (A) | 1.0A | 10 | Instrument, Medical Illuminator- Burning Position BUTH | CC-6 | 2 | 1.25 | 200 | 24 |
| 1813 | 14.4 | 27667 | T3.25 | Min Bay | | 50 | Indicator | C-2V | 1.19 | .62 | 1000 | |
| 1819 | 28 | 27711 | T3.25 | Min Bay | | 50 | Indicator | C-2F | 1.19 | .62 | 2500 | |
| 1820 | 28 | 27727 | T3.25 | Min Bay | | 50 | Indicator | C-2F | 1.19 | .62 | 1000 | |
| 1828 | 37.5 | 27772 | T3.25 | Min Bay | | 50 | | C-2F | 1.19 | .62 | 3000 | |
| 1829 | 28 | 27776 | T3.25 | Min Bay | | 50 | Indicator | C-2F | 1.19 | .62 | 1000 | |
| 1864 | 28 | 27862 | T3.25 | Min Bay | | 50 | | C-2F | 1.19 | .62 | 1500 | |
| 1876 | 3.5 | 27889 | T-5 | S.C. Bay. | 2.50A | 10 | Spec. Serv. Photoelectric Scan Silver plated contact (A.A. Tol. .016") | C-6 | 1.75 | 1.12 | 2000 | 6.5 |
| 1944 | 14 | 37034 | T-7 | S.C. Bay. | 3.57A | 10 | Special Service-Silver Contact | C-8Z | 2.25 | 1.25 | 300 | 75 |
| 1974 | 6 | 32780 | T-3 | WireTerm. | 20W | 10 | Instrument-Quartz bulb-Halogen 3" bare leads horizontal (304) | C-6 | 1.14 | 0.28 | 10,000+ | 10 |
| 2604x | 5 | 43805 | TL2.75 | G-4 2-Pin | 2A | 10 | Instrument - Halogen Lens end (306) | C-6 | 1.18 | - | 5000 | -(128) |
| 2331 | 5.9/ 6.2 | 28100 | RP-11 | D.C.Pf.(S) | 4.66A/ 4.49 | 10 | Instrument (305) | C-6 | 2.25 | 1.18 | 400 | 32 |
| 7387 | 28 | 28926 | T1.75 | Bi-Pin M-23 | | 50 | Indicator | C-2F | .61 | .5 | 7000 | |
| PR2/BP2 | 2.38 | 12675 | B3.5 | S.C. Min Flang | | 48 | Flashlight 2-D | C-2R | 1.25 | .25 | 15 | |
| TEL/24PSB | 24 | 12071 | T2 | Slide #5 | | 50 | Indicator | C-2F | 1.11 | | | |
| TEL/28PSB | 28 | 12072 | T2 | Slide #5 | | 50 | Telephone | C-2F | 1.11 | | | |
| TEL/120MB | 120 | 12078 | T2.5 | Min Bay | | 50 | Indicator | CC-7A | 1.19 | | | |
| TEL/120PSB | 120 | 12080 | T2 | Slide #5 | | 50 | Indicator | CC-7A | 1.11 | | | |

| Description | Volts | GE Product Code | Bulb | Base | Watts | Std. Pkg. Qty. | Additional Information | Filament Design | MOL (in.) | LCL (in.) | Rated Avg. Life hours | App. Initial Lumens |
|-------------------------------|-------|-----------------|------|-----------|-------|----------------|---|-----------------|-----------|-----------|-----------------------|---------------------|
| Specialty Incandescent | | | | | | | | | | | | |
| 3S6/5 | 130 | 11098 | S6 | Cand | 3 | 24 | Clear-Indicator | C-7A | 1.87 | 1.37 | 3000 | 11 |
| 15S11/14 | 120 | 13236 | S-11 | Cand. | 15 | 120 | Clear-Medical Spotlight | C-7A | 2.25 | 1.375 | 200 | 150 |
| 25S11/2C | 120 | 14585 | S-11 | Cand. | 25 | 120 | Clear-Medical Spotlight | C-7A | 2.25 | 1.437 | 500 | - |
| 150P25/2SB | 120 | 19334 | P-25 | Med.(BB) | | 150 | 60 Clear-Spotlight. Silvered Bowl. Hard glass button (28) | C-7A | 4.75 | - | 200 | - |
| 150P25/10 | 120 | 19372 | P-25 | Med. (BB) | | 150 | 60 Light I.F. - Spotlight Hard glass button | C-5 | 4.75 | 3 | 200 | 2100 |

Specialty Fluorescent

| | | | | | | | | | | | |
|------------|-------|-----|------------------|----|----|--|---|------|---|--------|---|
| F4T5/BL | 10017 | T5 | Min Bi-Pin (G5) | 4 | 24 | Blacklight, UVA Source (8) | - | 6 | - | 5000 | - |
| F4T5/BLB | 10019 | T5 | Min Bi-pin (G5) | 4 | 24 | Blacklight Blue, UVA Source, Integral Dark Blue Filter | - | 6 | - | 5000 | - |
| F8T5/BLB | 10077 | T5 | Min Bi-pin (G5) | 8 | 24 | Blacklight Blue, UVA Source, Integral Dark Blue Filter | - | 12 | - | 5000 | - |
| F15T8/BL | 35884 | T8 | Med Bi-pin (G13) | 15 | 24 | Blacklight, UVA Source (8) | - | 18 | - | 7500 | - |
| F15T8/BLB | 35885 | T8 | Med Bi-pin (G13) | 15 | 24 | Blacklight Blue, UVA Source, Integral Dark Blue Filter | - | 18 | - | 7500 | - |
| F20T12/B | 10231 | T12 | Med Bi-Pin (G13) | 20 | 24 | Blue | - | 24 | - | 9000 | - |
| F20T12/BL | 10244 | T12 | Med Bi-Pin (G13) | 20 | 24 | Blacklight, UVA Source (8) | - | 24 | - | 9000 | - |
| F20T12/BLB | 34747 | T12 | Med Bi-Pin (G13) | 20 | 24 | Blacklight Blue, UVA Source, Integral Dark Blue Filter | - | 24 | - | 9000 | - |
| F40B | 10514 | T12 | Med Bi-Pin (G13) | 40 | 24 | Blue | - | 48 | - | 20,000 | - |
| F40BL | 10526 | T12 | Med Bi-Pin (G13) | 40 | 24 | Blacklight, UVA Source (8) | - | 48 | - | 20,000 | - |
| F40BL/U/3 | 40537 | T12 | Med Bi-Pin (G13) | 40 | 12 | Blacklight, UVA Source, Mod-U-Line® 3.625 Spacing Between Legs | - | 22.5 | - | 14,000 | - |
| F40BLB | 10531 | T12 | Med Bi-Pin (G13) | 40 | 24 | Blacklight Blue, UVA Source, Integral Dark Blue Filter | - | 48 | - | 20,000 | - |
| G8T5 | 11077 | T5 | Min Bi-Pin (G15) | 8 | 24 | Clear, Preheat, for use with Starters UVC Source (9) | - | 12 | - | 7500 | - |
| G64T6 | 11086 | T6 | Single Pin (Fa8) | 65 | 24 | Clear, Instant Start 425mA (9) | - | 64 | - | 7500 | - |
| G15T8 | 11078 | T8 | Med Bi-Pin (G13) | 15 | 24 | Clear, Preheat, for use with Starters UVC Source (9) | - | 18 | - | 7500 | - |
| G25T8 | 11082 | T8 | Med Bi-Pin (G13) | 25 | 24 | Clear, Preheat, for use with Starters UVC Source (9) | - | 18 | - | 7500 | - |
| G30T8 | 11080 | T8 | Med Bi-Pin (G13) | 30 | 24 | Clear, Preheat, for use with Starters UVC Source (9) | - | 36 | - | 7500 | - |

- (8) CAUTION: Risk Group 1 (Low Risk): UV emitted from this lamp. Skin or eye irritation could result. Minimize exposure.
- (9) Warning: Risk Group 3 (High Risk): UV emitted from this lamp. Avoid exposure of eyes and skin to unshielded lamp. Skin or eye injury will result.
- (28) For use only in porcelain sockets and fixtures so designed that the temperatures of the lamp and fixture do not exceed limits for satisfactory operation.
- (73) CAUTION: While this lamp is carefully inspected before shipment, the glass bulb may crack when subjected to abnormal pressure. Therefore, it is recommended that the bulb be grasped with a cloth or glove when removing or installing the lamp in a tight fitting socket
- (101) Made in Japan expressly for General Electric Company
- (103) Life dependent on service conditions
- (114) CAUTION: This halogen-cycle bulb could shatter if scratched or damaged Use appropriate protection when handling, using or disposing. Use in fixtures designed for the high temperature required for proper operation and that offer protection in case the bulb shatters. Turn power off when changing lamps. Allow lamp to cool before removal. For satisfactory performance: (1) limit seal and outer lead wire temperature to 350°C or lead wire deterioration may occur; (2) maintain a minimum bulb wall temperature of 250°C for operation of the halogen cycle; (3) remove grease or fingerprints from bulb by cleaning with a grease-free solvent; (4) operate at design volts

(304) Warning

Risk of Fire

- Keep combustible materials away from lamp
- Use in fixture rated for this product

Lamp emits UV radiation which may cause eye/skin injury

- Avoid exposure of eyes and skin to unshielded lamp

Pressurized lamp-unexpected rupture may cause injury, fire, or property damage

- Use sys protection when handling lamp
- Do not exceed rated voltage
- Do not touch glass with bare hands
- Do not use lamp if outer glass is scratched or broken
- Dispose of lamp in a closed container

Do not turn on lamp until fully installed

- Keep away from children
- Use protective screen when handling

Caution

Risk of burn

- Allow lamp/fixture to cool before handling

FOR BEST PERFORMANCE

- Limit seal temperature to 350° C
- Maintain 250° C minimum bulb wall temperature
- Remove fingerprints from bulb with grease-free solvent
- Operate at design voltage

(305) CAUTION

- Lamp may shatter and cause injury if broken**
- Do not use excessive force when installing lamp

(306) WARNING

Pressurized lamp - unexpected rupture may cause injury, fire, or property damage

- Use eye protection when handling lamp
- Do not exceed rated voltage
- Avoid direct water/liquid contact
- Use in enclosed fixture rated for this product
- Do not use lamp if outer glass is scratched or broken
- Dispose of lamp in a closed container
- Do not turn on lamp until fully installed
- Keep away form children
- Use protective screen when handling

CAUTION

Risk of burn

- Allow lamp/fixture to cool before handling

勝特力材料 886-3-5753170
 勝特力电子(上海) 86-21-54151736
 勝特力电子(深圳) 86-755-83298787
[Http://www.100y.com.tw](http://www.100y.com.tw)



GE Lighting